


Chromatograph Calibration Table

Max Divisions	Max Volts	Constant	Column Pressure	Total Gas Panel Test	2% Methane Panel Read	20% Methane Panel Read
2000	5.000	400	3 lbs	Panel Tested OK	100 UNITS	1000

Test Date	1/19/2013
-----------	-----------

CHROMATOGRAPH



Test Gas 1%	Volts Read During Test	Actual Divisions Read	Percent Per Division
C1	0.096	38	0.02604
C2	0.094	38	0.02660
C3	0.063	25	0.03968
IC4	0.044	18	0.05682
NC4	0.036	14	0.06944

Enter Your Voltage Readings Below To Calculate Max % of Gases

Type Gas	Max Voltage Read In Show	Divisions From Show	Max Gas In Percent	Enter Max Gas Units	Gas % Per Unit/Div. Ratios
C1	0.096	38.4	1.000	300	0.00333
C2	0.094	37.6	1.000		0.00333
C3	0.063	25.2	1.000		0.00333
IC4	0.044	17.6	1.000		0.00333
NC4	0.036	14.4	1.000		0.00333

Unit Check List

ITEM	CONDITION	ITEM	CONDITION	ITEM	CONDITION
Computer	Yes	Vacuum Check	Yes		
Sensor Panel	Yes	Sample Pump	Yes		
WITS Box	Yes	Polyflow Lines	Yes		
A/C	Yes	BlowBack Valve	Yes		
Total Gas	Yes	Labels In Place	Yes		
Chromatograph	Yes	Breakers Off	Yes		
Printer	Yes	Lights Work	Yes		
Air Compressor	Yes	Drawers	Yes		